

# DECLARATION OF CONFORMITY



Application of Council Directive: 2004/108/EEC – The EMC Directive  
2006/95/EEC – The Low Voltage Directive  
2012/19/EC – WEEE Directive

Manufacturer's Name: Teledyne Isco, Inc.  
Manufacturer's Address: 4700 Superior, Lincoln, Nebraska 68504 USA  
Mailing Address: P.O. Box 82531, Lincoln, NE 68501

Equipment Type/Environment: Laboratory Equipment for Light Industrial/Commercial Environments  
Trade Name/Model No: GLS Sampler with 923 power supply  
Year of Issue: 2013

Standards to which Conformity is Declared: EN 61326-1:2006 EMC Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use  
EN 61010-1:2010 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use

Standard	Description	Severity Applied	Performance Criteria
EN61000-4-2: 2009	Electrostatic Discharge	Level 2 - 4kV contact discharge Level 3 - 8kV air discharge	A A
EN61000-4-3: 2010	Radiated RF Immunity	80 MHz to 2.7GHz 80% AM at 1kHz Level 1 - 10V/m	A
EN61000-4-4: 2010	Electrical Fast Transient on Mains and I/O	Level 2 - 2kV on ac lines	A
EN61000-4-5: 2005	Surge on AC Lines	2kV common mode, 1kV differential mode	A
EN61000-4-6: 2008	Conducted RF on AC Mains and I/O	150 kHz to 80 MHz, 3V rms, 80% modulated	A
EN61000-4-11: 2004	Voltage Dips/Short Interruptions	0% during 1 cycle and full cycle 70% at 25 cycles	A
CISPR11/ EN 55011: 2009	RF Emissions Radiated, below 1GHz and Conducted, AC Mains	Group 1, Class A Industrial, Scientific, and Medical Equipment	PASS
EN61000-3-2: 2009, 3-3: 2008	AC Harmonic, Flicker		PASS

We, the undersigned, hereby declare that the design of the equipment specified above conforms to the above Directive(s) and Standards as of August 27, 2013.

USA Representative

**Vikas V. Padhye Ph, D**  
Vice President and Product General Manager  
Teledyne Isco, Inc.  
4700 Superior Street  
Lincoln, Nebraska 68504  
Phone: (402) 464-0231  
Fax: (402) 464-0318



60-2953-041  
Rev D