

DECLARATION OF CONFORMITY



Application of Council Directive: 89/336/EEC – The EMC Directive
72/23/EEC – The Low Voltage Directive

Manufacturer's Name: Insite IG
Manufacturer's Address: 80 Whisperwood Blvd. Ste 107,
Slidell LA

Equipment Type/Environment: Measurement & Process Control/Industrial
Trade Name/Model No: Process Analyzer/Model 2000
Year of Issue: 2004

Standards to which Conformity is Declared: EN 50081-2 (EN55011 Group 1 Class A)
EN 61000-4
EN 61010B-1

Test Standard	Test Specification	Performance	Results
61000-4-2:00, ESD	±8kV Air Discharge	B	C
61000-4-3:00, Radiated RF	3V/m, 2s, 80MHz, 80% AM 1kHz rate	A	A
61000-4-4:00, EFT/Burst	±2kVLL, ±1kV Signal Port	A	B
61000-4-5:00, Surge	±1kV LL, ±2kV LG	B	A
61000-4-6:00, Conducted RF	3Vrms, 2s, 150kHz-30MHz, 80%AM 1kHz rate	A	B
61000-4-11:00, Voltage Dip	100% drop, 20ms, 0deg and 180deg	B	A

We, the undersigned, hereby declare that the design of the equipment specified above conforms to the above Directives and Standards and of March 2004.

Richard Davis
Vice President
Insite Instrumentation Group
Slidell, LA USA